

## 1 Instruments

The firmware version below can be used with the following instruments:

- 930 Compact IC Flex
- 940 Professional IC
- 943 Professional Thermostat/Reactor Vario
- 945 Professional Detector Vario

## 2 Purpose

This document describes the new features, the improvements and the fixed bugs in the following new firmware version:

- 5.940.0102

In comparison with the firmware version:

- 5.940.0101

## 3 New features

Topic	Description
2.850.9020 IC Conductivity Detector MB	The 2.850.9020 IC Conductivity Detector MB is now supported.

## 4 Improvements

Topic	Description
None	None

## 5 Fixed bugs

Topic	Description
None	None

## 6 Known bugs

Topic	Description
None	None

## 7 Conformity

The firmware version 5.940.0102 does not contain any modifications that affect conformity of the instruments mentioned above regarding 21 CFR Part 11 and GAMP.