



## 946 Portable VA Analyzer

### General information

---

Software version: 1.0

Instrument: -

Sensor: scTRACE Gold

User name:

Report: No

Report elements: -

## Method

### General

---

Method name: AB422 Cleaning scTRACE Gold.detp

Remarks: 15 mL ultrapure water, 1.5 mL electrolyte (3% formic acid, 60 mmol/L KCl, 5 mg/L Fe)

### Determination

---

Sample volume (mL): 16.5

Total cell volume (mL): 16.5

Stirring time (s): 10.0

Stirring rate (1/min): 2000

Measure blank: No

No. of blanks: 0

Blank value correction: No

No. of replications: 0

No. of additions: 0

### Voltammetric

---

Measuring mode: Differential pulse

Current measuring range: +/- 10 uA

#### Cyclovoltammetric pretreatment

Start potential (V): -0.7

Vertex potential (V): 1.0

Potential step (V): 0.01

Sweep rate (V/s): 1.0

No. of cycles: 20

#### Potentiostatic pretreatment

Potential 1 (V): 0.9

Waiting time 1 (s): 10.0

Potential 2 (V): 0.3

Waiting time 2 (s): 5.0

Equilibration time (s): 5.0

#### Sweep

Start potential (V): 0.3

End potential (V): 0.6

Potential step (V): 0.004



## 946 Portable VA Analyzer

Pulse amplitude (V): 0.05

Pulse time (s): 0.04

Sweep rate (V/s): 0.01

Evaluation

Data processing

Smoothing: 1  
Calibration method: Standard addition

Peak evaluation

	-
Characteristic potential (V)	0.0
Tolerance (V)	0.05
Min. width (V)	0.05
Max. width (V)	0.5
Min. measured quantity (µA)	0.001
Baseline type	Linear
Base point automatic	Yes
Start base point (V)	0.0
End base point (V)	0.0

Standard solutions

	-	Volume (mL)
Standard 1	1.0 mg/L	0.1
Standard 2	-	-
Standard 3	-	-
Standard 4	-	-