Method run

Method name: AB445 - Maintenance - Cleaning measurement
Method version: 1
Method group: Main group

START — Main track

CALL — Empty measuring vessel — Drain / Drain

LQH — Dose cleaning solution

CALL VA — Clean electrode — Clean electrode / VA track

CALL — CALL Rinsing — Rinsing cycle / Rinsing cycle

END — Main track

VA TRACK — VA track

LOOP — Replications

DP — Sweep

END — VA track

TRACK — Rinsing cycle

LOOP — LOOP Rinsing

CALL — CALL Drain — Drain / Drain

CALL — CALL Rinse — Rinse / Rinse

END — Rinsing cycle

TRACK — Drain

PUMP — PUMP Drain

END — Drain

TRACK — Rinse

PUMP — PUMP Rinse

END — Rinse

TRACK — Shut off

PUMP — PUMPS OFF

STIR & PURGE — STIR & PURGE OFF

LQH — Fill DU electrolyte

END — Shut off

EXIT — Exit track

CALL — Exit shut off — Exit shut off / Shut off

END — Exit track

ERROR — Error track

CALL — Error shut off — Error shut off / Shut off

END — Error track