



dAg Titrode

6.00404.300

Digital, combined silver ring electrode for OMNIS with a pH glass membrane as reference electrode.

This maintenance-free electrode is suitable for precipitation titrations where the pH value remains constant (titrant silver nitrate), e.g. of

- Chloride, Bromide, Iodide
- Sulfides
- Hydrogen sulfide
- Mercaptans
- Cyanides

This electrode is stored in distilled water.
dTrodes can be used on OMNIS Titrators.

Below, the accessories are grouped into Scope of delivery and Optional accessories.
Please keep this printout at hand for ordering replacement material.
These lists may be subject to change.

Scope of delivery 6.00404.300

Qt.	Order no.	Description
1 PCS	6.1236.050	Sleeve with SGJ 14/12 mm Sleeve with SGJ 14/12 mm, polyethylene.
		
1 PCS	6.2008.040	Storage vessel Together with 6.2008.050 storage vessel holder. it provides a support for the electrode on 807 Dosing Units.
		

Optional accessories

Order no.	Description	
6.02104.300	Electrode cable plug-in head Q / plug P, 0.55 m Cable for connecting a dTrobe (plug-in head Q) to the digital measuring module from OMNIS.	
6.02104.310	Electrode cable plug-in head Q / plug P, 1.5 m Cable for connecting a dTrobe (plug-in head Q) to the digital measuring module from OMNIS.	