Electrodo Ionic Surfactant

6.0507.120

Este electrodo de tensioactivos debe utilizarse en combinación con un electrodo de referencia y es adecuado para, por ejemplo:

- Titulación de tensioactivos aniónicos en matrices acuosas
- Titulación de tensioactivos catiónicos en matrices acuosas

El electrodo muestra un excelente comportamiento de respuesta debido a los ionóforos inmovilizados en la membrana.

Partes/accesorios 6.0507.120

Qt.	Order no.	Descripción

1 PCS 6.1236.020 Sleeve with SGJ 14/12 mm

Sleeve with SGJ 14/12 mm and O-ring.

---

1 PCS 6.2008.040 Storage vessel

Together with 6.2008.050 storage vessel holder. it provides a support for the electrode on 807 Dosing Units.

Accesorios opcionales

Order no.	Descripción
6.0726.100	Ag/AgCl reference electrode (length 12.5 cm) Silver/silver chloride reference electrode with Double-Junction System, installation length 10 cm. The standard ground-joint 14/15 enables easy assembly, and the flexible ground-joint diaphragm, which is insensitive to contamination, can be replaced at any time. The reference electrolyte and the bridge electrolyte can be selected freely according to use, and are easy to replace. This sensor is supplied without electrolyte filling. 
6.0750.100	LL ISE reference electrode Silver / silver chloride reference electrode with double junction system. This reference electrode is well suited for: • automated applications • ion measurements • surfactant titrations The ground-joint diaphragm, which is insensitive to contamination, offers a constant and reproducible electrolyte outflow. Additionally, the reference electrolyte is gelified for even better signal stability. The sensor is delivered with c(KCl) = 3 mol/L as bridge electrolyte, which can be freely selected and exchanged as needed. 
6.2104.020	Electrode cable / 1 m / F For connecting electrodes with Metrohm plug-in head G to Metrohm instruments (socket F).