



Titrode Ag

6.0430.100

Électrode annulaire en argent combinée à membrane de verre pH comme électrode de référence. Cette électrode ne nécessitant aucune maintenance convient aux titrages par précipitation en cas de valeur pH constante (nitrate d'argent comme titrant), par ex. de :

- Chlorure, bromure, iodure
- Sulfures
- Sulfure d'hydrogène
- Mercaptans
- Cyanures

Cette électrode est conservée dans l'eau distillée.

En fonction de l'utilisation, il est recommandé d'utiliser une titrode en argent avec revêtement (d'AgBr, AgCl ou Ag₂S), qui peut être commandée en conséquence.

Contenu de la livraison 6.0430.100

Qt.	Order no.	Description
-----	-----------	-------------

1 PCS

6.1236.050

Sleeve with SGJ 14/12 mm

Sleeve with SGJ 14/12 mm, polyethylene.



1 PCS

6.2008.040

Storage vessel

Together with 6.2008.050 storage vessel holder, it provides a support for the electrode on 807 Dosing Units.



Accessoires optionnels

Order no.	Description
6.2104.020	<p>Electrode cable / 1 m / F</p> <p>For connecting electrodes with Metrohm plug-in head G to Metrohm instruments (socket F).</p> 